

		<h1>ESD CERTIFICATE</h1>	
Originator: J.F. ROBLES	Revision N°: 002	Revision date:	21.04.2022
Test Laboratory	Centexbel GRÂCE-HOLLOGNE BE-4460 Grâce-Hollogne, Belgium		
Test N°	18.01970.01		

SHIELDskin XTREME™ Sterile ORANGE NITRILE™ 300 DI

Code: 69 655X – Orange Nitrile Gloves, sterile, gamma irradiated

SURFACE RESISTIVITY

Reference T1807934 sample tested as received

Surface Resistivity Information in accordance with EN 1149-1:2006	
Measured Ohms	$3.28 \times 10^{12} \Omega$

Air temperature = 23°C ± 1 / Relative humidity = 25 % ± 2 %

VERTICAL RESISTANCE

Reference T1807934 sample tested as received

Vertical Resistance Information in accordance with EN 1149-2:1997	
Measured Ohms	$9.64 \times 10^9 \Omega$

Air temperature = 23°C ± 1 / Relative humidity = 25 % ± 2 %

DECAY CHARGE

Reference T1807934 sample tested as received

Dimension of the specimens: sample too small. Measurements made with the little ring.

Decay Charge Information in accordance with EN 1149-3:2004	
Shielding Factor S	0.00
Measured Half decay time T50 (S)	1.95

Air temperature = 23°C ± 1 / Relative humidity = 25 % ± 2 %

SHIELD Scientific B.V.



J.F. ROBLES
General Manager

The data provided is based on gloves tested under laboratory conditions, in accordance with EN1149-1:2006, EN1149-2:1997 and EN1149-3:2004. The information is for guidance only and may not reflect the user's application. A risk assessment should always be made by purchaser to assess the suitability of gloves for a specific application.